

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01767	Calib. Date: 05/01/2019
Certification File No.:	01768	Calib. No.: 00042
Linearity File No.:	01769	Cert. Date: 05/01/2019
Solution File No.:	01779	Cert. No.: 00032
Sequential File No.:	01779	Lin. Date: 05/01/2019
		Lin. No.: 00032
		Soln. Date: 08/01/2019
		Soln. No.: 00288
		File Date: 08/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 0791

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:19D		
Control 1 EC	0.100%	12:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 2 EC	0.099%	12:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		
Control 3 EC	0.099%	12:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 08/01/2019