

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02617	Calib. Date: 05/01/2019
Certification File No.:	02618	Calib. No.: 00033
Linearity File No.:	02619	Cert. Date: 05/01/2019
Solution File No.:	02689	Cert. No.: 00024
Sequential File No.:	02689	Lin. Date: 05/01/2019
		Lin. No.: 00024
		Soln. Date: 08/11/2019
		Soln. No.: 00327
		File Date: 08/11/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %: 0.100%		Expires: 09/26/2020
Solution Control Lot: 18320		Bottle No.: 1146

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:04D		
Control 1 EC	0.098%	09:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05D		
Control 2 EC	0.097%	09:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07D		
Control 3 EC	0.097%	09:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:08D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BARTOSHEK

First Name: DAVID

MI: G

Signature: 

Badge No.: 316

Date: 08/11/2019