

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01767 Calib. Date: 05/01/2019 Calib. No.: 00042
Certification File No.: 01768 Cert. Date: 05/01/2019 Cert. No.: 00032
Linearity File No.: 01769 Lin. Date: 05/01/2019 Lin. No.: 00032
Solution File No.: 01784 Soln. Date: 08/23/2019 Soln. No.: 00289
Sequential File No.: 01784 File Date: 08/23/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 0069

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:36D		
Control 1 EC	0.099%	14:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		
Control 2 EC	0.098%	14:38D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		
Control 3 EC	0.097%	14:39D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 08/23/2019