

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02617 Calib. Date: 05/01/2019 Calib. No.: 00033
Certification File No.: 02618 Cert. Date: 05/01/2019 Cert. No.: 00024
Linearity File No.: 02619 Lin. Date: 05/01/2019 Lin. No.: 00024
Solution File No.: 02707 Soln. Date: 08/27/2019 Soln. No.: 00329
Sequential File No.: 02707 File Date: 08/27/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 0727

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:50D		
Control 1 EC	0.099%	20:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:51D		
Control 2 EC	0.098%	20:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:52D		
Control 3 EC	0.098%	20:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BARTOSHEK

First Name: DAVID

MI: G

Signature: _____

David H. Bartoshek #316 TCRD

Badge No.: 316

Date: 08/27/2019