

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 02617  
 Certification File No.: 02618  
 Linearity File No.: 02619  
 Solution File No.: 02712  
 Sequential File No.: 02712

Serial No.: ARAJ-0075

Calib. Date: 05/01/2019 Calib. No.: 00033  
 Cert. Date: 05/01/2019 Cert. No.: 00024  
 Lin. Date: 05/01/2019 Lin. No.: 00024  
 Soln. Date: 09/02/2019 Soln. No.: 00330  
 File Date: 09/02/2019

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 19060

Model No.: CU-34

Serial No.: DDAN-0041  
 Expires: 02/11/2021  
 Bottle No.: 0147

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:56D		
Control 1 EC	0.101%	02:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:57D		
Control 2 EC	0.099%	02:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:59D		
Control 3 EC	0.098%	02:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_



Badge No.: 365

Date: 09/02/2019