

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01793 Calib. Date: 09/09/2019 Calib. No.: 00043  
Certification File No.: 01794 Cert. Date: 09/09/2019 Cert. No.: 00033  
Linearity File No.: 01795 Lin. Date: 09/09/2019 Lin. No.: 00033  
Solution File No.: 01796 Soln. Date: 09/09/2019 Soln. No.: 00290  
Sequential File No.: 01796 File Date: 09/09/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387  
Control Solution %: 0.100% Expires: 02/11/2021  
Solution Control Lot: 19060 Bottle No.: 0010

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:30D		
Control 1 EC	0.100%	14:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		
Control 2 EC	0.099%	14:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33D		
Control 3 EC	0.099%	14:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-026 (DL)

**Changed By:**

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpc I Dent 7045

Badge No.: 7045

Date: 09/09/2019