

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02717	Calib. Date: 09/09/2019
Certification File No.:	02718	Calib. No.: 00034
Linearity File No.:	02719	Cert. Date: 09/09/2019
Solution File No.:	02720	Cert. No.: 00025
Sequential File No.:	02720	Lin. Date: 09/09/2019
		Lin. No.: 00025
		Soln. Date: 09/09/2019
		Soln. No.: 00331
		File Date: 09/09/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 0066

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:23D		
Control 1 EC	0.100%	11:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 2 EC	0.099%	11:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 3 EC	0.100%	11:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-025 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: _____

TP I [Signature] 7045

Badge No.: 7045

Date: 09/09/2019