

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01793 Calib. Date: 09/09/2019 Calib. No.: 00043
Certification File No.: 01794 Cert. Date: 09/09/2019 Cert. No.: 00033
Linearity File No.: 01795 Lin. Date: 09/09/2019 Lin. No.: 00033
Solution File No.: 01799 Soln. Date: 09/28/2019 Soln. No.: 00291
Sequential File No.: 01799 File Date: 09/28/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 0064

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:09D		
Control 1 EC	0.099%	19:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:11D		
Control 2 EC	0.099%	19:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:12D		
Control 3 EC	0.099%	19:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 09/28/2019