

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02717 Calib. Date: 09/09/2019 Calib. No.: 00034  
Certification File No.: 02718 Cert. Date: 09/09/2019 Cert. No.: 00025  
Linearity File No.: 02719 Lin. Date: 09/09/2019 Lin. No.: 00025  
Solution File No.: 02731 Soln. Date: 09/27/2019 Soln. No.: 00333  
Sequential File No.: 02731 File Date: 09/27/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 02/11/2021  
Solution Control Lot: 19060 Bottle No.: 0016

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:33D		
Control 1 EC	0.099%	17:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:34D		
Control 2 EC	0.098%	17:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36D		
Control 3 EC	0.098%	17:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WORTH

First Name: ADAM

MI: C

Signature:



Badge No.: 378

Date: 09/27/2019