

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01793	Calib. Date:	09/09/2019
Certification File No.:	01794	Cert. Date:	09/09/2019
Linearity File No.:	01795	Lin. Date:	09/09/2019
Solution File No.:	01802	Soln. Date:	10/19/2019
Sequential File No.:	01802	File Date:	10/19/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0150

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:08D		
Control 1 EC	0.100%	12:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09D		
Control 2 EC	0.100%	12:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10D		
Control 3 EC	0.099%	12:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 10/19/2019