

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02717	Calib. Date:	09/09/2019
Certification File No.:	02718	Cert. No.:	00034
Linearity File No.:	02719	Cert. Date:	09/09/2019
Solution File No.:	02773	Lin. Date:	09/09/2019
Sequential File No.:	02773	Lin. No.:	00025
		Soln. Date:	11/09/2019
		Soln. No.:	00340
		File Date:	11/09/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0206


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:55S		
Control 1 EC	0.102%	22:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	22:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:56S		
Control 2 EC	0.100%	22:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:58S		
Control 3 EC	0.099%	22:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	22:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature:  # 365 Badge No.: 365
 Date: 11/09/2019