

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01793 Calib. Date: 09/09/2019 Calib. No.: 00043
Certification File No.: 01794 Cert. Date: 09/09/2019 Cert. No.: 00033
Linearity File No.: 01795 Lin. Date: 09/09/2019 Lin. No.: 00033
Solution File No.: 01811 Soln. Date: 11/14/2019 Soln. No.: 00293
Sequential File No.: 01811 File Date: 11/14/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0044

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:15S		
Control 1 EC	0.100%	12:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16S		
Control 2 EC	0.100%	12:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17S		
Control 3 EC	0.100%	12:18S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 11/14/2019