

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02717	Calib. Date: 09/09/2019
Certification File No.:	02718	Calib. No.: 00034
Linearity File No.:	02719	Cert. Date: 09/09/2019
Solution File No.:	02785	Cert. No.: 00025
Sequential File No.:	02785	Lin. Date: 09/09/2019
		Lin. No.: 00025
		Soln. Date: 11/20/2019
		Soln. No.: 00341
		File Date: 11/20/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 0047

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:54S		
Control 1 EC	0.102%	12:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55S		
Control 2 EC	0.100%	12:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56S		
Control 3 EC	0.099%	12:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA	First Name: PASCAL	MI: M
Signature: <u> #365</u>		Badge No.: 365
		Date: 11/20/2019