

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01793	Calib. Date: 09/09/2019
Certification File No.:	01794	Calib. No.: 00043
Linearity File No.:	01795	Cert. Date: 09/09/2019
Solution File No.:	01814	Cert. No.: 00033
Sequential File No.:	01814	Lin. Date: 09/09/2019
		Lin. No.: 00033
		Soln. Date: 12/06/2019
		Soln. No.: 00294
		File Date: 12/06/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 0014

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:26S		
Control 1 EC	0.102%	15:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		
Control 2 EC	0.099%	15:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29S		
Control 3 EC	0.099%	15:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 12/06/2019