

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01793	Calib. Date: 09/09/2019
Certification File No.:	01794	Calib. No.: 00043
Linearity File No.:	01795	Cert. Date: 09/09/2019
Solution File No.:	01816	Cert. No.: 00033
Sequential File No.:	01816	Lin. Date: 09/09/2019
		Lin. No.: 00033
		Soln. Date: 12/27/2019
		Soln. No.: 00295
		File Date: 12/27/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 0392

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:10S		
Control 1 EC	0.100%	10:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12S		
Control 2 EC	0.100%	10:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		
Control 3 EC	0.100%	10:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA	First Name: PASCAL	MI: M
Signature: <u> #365</u>	Badge No.: 365	
	Date: 12/27/2019	