

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02717 Calib. Date: 09/09/2019 Calib. No.: 00034
Certification File No.: 02718 Cert. Date: 09/09/2019 Cert. No.: 00025
Linearity File No.: 02719 Lin. Date: 09/09/2019 Lin. No.: 00025
Solution File No.: 02821 Soln. Date: 01/11/2020 Soln. No.: 00346
Sequential File No.: 02821 File Date: 01/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0040

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:17S		
Control 1 EC	0.100%	11:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 2 EC	0.100%	11:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		
Control 3 EC	0.098%	11:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 01/11/2020