

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	09/09/2019
Calibration File No.:	02717	Calib. No.:	00034
Certification File No.:	02718	Cert. Date:	09/09/2019
Linearity File No.:	02719	Lin. Date:	09/09/2019
Solution File No.:	02826	Soln. Date:	01/19/2020
Sequential File No.:	02826	File Date:	01/19/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0452

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:13S		
Control 1 EC	0.101%	04:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15S		
Control 2 EC	0.098%	04:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16S		
Control 3 EC	0.098%	04:17S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 01/19/2020