

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01793	Calib. Date: 09/09/2019
Certification File No.:	01794	Calib. No.: 00043
Linearity File No.:	01795	Cert. Date: 09/09/2019
Solution File No.:	01819	Cert. No.: 00033
Sequential File No.:	01819	Lin. Date: 09/09/2019
		Lin. No.: 00033
		Soln. Date: 01/25/2020
		Soln. No.: 00296
		File Date: 01/25/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWK S3-0387
Control Solution %:	0.100%		Expires: 09/09/2021
Solution Control Lot:	19240		Bottle No.: 0220

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:56S		
Control 1 EC	0.100%	13:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:57S		
Control 2 EC	0.099%	13:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:59S		
Control 3 EC	0.099%	13:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 01/25/2020