

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02717	Calib. Date: 09/09/2019
Certification File No.:	02718	Calib. No.: 00034
Linearity File No.:	02719	Cert. Date: 09/09/2019
Solution File No.:	02834	Cert. No.: 00025
Sequential File No.:	02834	Lin. Date: 09/09/2019
		Lin. No.: 00025
		Soln. Date: 02/06/2020
		Soln. No.: 00348
		File Date: 02/06/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 0098

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:35S		
Control 1 EC	0.103%	14:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:36S		
Control 2 EC	0.100%	14:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37S		
Control 3 EC	0.100%	14:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 02/06/2020