

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 02843  
 Certification File No.: 02844  
 Linearity File No.: 02845  
 Solution File No.: 02846  
 Sequential File No.: 02846

Serial No.: ARAJ-0075

Calib. Date: 02/14/2020      Calib. No.: 00035  
 Cert. Date: 02/14/2020      Cert. No.: 00026  
 Lin. Date: 02/14/2020      Lin. No.: 00026  
 Soln. Date: 02/14/2020      Soln. No.: 00349  
 File Date: 02/14/2020

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 19240

Model No.: CU-34

Serial No.: DDAN-0041  
 Expires: 09/09/2021  
 Bottle No.: 0274

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:19S		
Control 1 EC	0.099%	10:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		
Control 2 EC	0.096%	10:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22S		
Control 3 EC	0.096%	10:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436 (DL)

**Changed By:**

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I Lutz 7045

Badge No.: 7045

Date: 02/14/2020