

SOLN 298

2/18/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01826	Calib. Date:	02/18/2020
Certification File No.:	01827	Cert. No.:	00045
Linearity File No.:	01828	Cert. Date:	02/18/2020
Solution File No.:	01829	Lin. Date:	02/18/2020
Sequential File No.:	01829	Lin. No.:	00035
		Soln. Date:	02/18/2020
		Soln. No.:	00298
		File Date:	02/18/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0097

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:06S		
Control 1 EC	0.099%	15:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07S		
Control 2 EC	0.098%	15:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09S		
Control 3 EC	0.098%	15:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXK P2-444 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I [Signature] 7045

Badge No.: 7045

Date: 02/18/2020