

SOLN 299  
2/19/2020

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01826	Calib. Date:	02/18/2020
Certification File No.:	01827	Cert. No.:	00045
Linearity File No.:	01828	Cert. Date:	02/18/2020
Solution File No.:	01831	Lin. Date:	02/18/2020
Sequential File No.:	01831	Lin. No.:	00035
		Soln. Date:	02/19/2020
		Soln. No.:	00299
		File Date:	02/19/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0218

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:05S		
Control 1 EC	0.100%	10:05S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		
Control 2 EC	0.097%	10:07S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08S		
Control 3 EC	0.098%	10:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHWARTZ First Name: STEVE

MI: -

Signature: 

Badge No.: 1293

Date: 02/19/2020