

SOLN 300

2/25/2020

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
 Location: TOMS RIVER P.D.
 Calibration File No.: 01826 Calib. Date: 02/18/2020 Calib. No.: 00045
 Certification File No.: 01827 Cert. Date: 02/18/2020 Cert. No.: 00035
 Linearity File No.: 01828 Lin. Date: 02/18/2020 Lin. No.: 00035
 Solution File No.: 01837 Soln. Date: 02/25/2020 Soln. No.: 00300
 Sequential File No.: 01837 File Date: 02/25/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 0163

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:55S		
Control 1 EC	0.097%	10:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 2 EC	0.097%	10:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58S		
Control 3 EC	0.097%	10:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00S		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHWARTZ First Name: STEVE

MI: -

Signature: 

Badge No.: 1293
Date: 02/25/2020