

SOLN 352

2/25/2020

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02843  
Certification File No.: 02844  
Linearity File No.: 02845  
Solution File No.: 02855  
Sequential File No.: 02855

Serial No.: ARAJ-0075

Calib. Date: 02/14/2020  
Cert. Date: 02/14/2020  
Lin. Date: 02/14/2020  
Soln. Date: 02/25/2020  
File Date: 02/25/2020

Calib. No.: 00035  
Cert. No.: 00026  
Lin. No.: 00026  
Soln. No.: 00352

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 19240

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 09/09/2021  
Bottle No.: 0162

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:11S		
Control 1 EC	0.099%	11:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 2 EC	0.096%	11:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		
Control 3 EC	0.096%	11:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: \_\_\_\_\_



Badge No.: 1293

Date: 02/25/2020