

SOLN. 355

4/18/2020

Calibrating Unit New Standard Solution Report


Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
 Location: TOMS RIVER P.D.
 Calibration File No.: 02859 Calib. Date: 04/03/2020 Calib. No.: 00037
 Certification File No.: 02860 Cert. Date: 04/03/2020 Cert. No.: 00027
 Linearity File No.: 02861 Lin. Date: 04/03/2020 Lin. No.: 00027
 Solution File No.: 02869 Soln. Date: 04/18/2020 Soln. No.: 00355
 Sequential File No.: 02869 File Date: 04/18/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
 Control Solution %: 0.100% Expires: 09/09/2021
 Solution Control Lot: 19240 Bottle No.: 0042

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:26D		
Control 1 EC	0.100%	13:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:28D		
Control 2 EC	0.100%	13:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29D		
Control 3 EC	0.100%	13:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: GAMBARDELLA First Name: PASCAL MI: M
 Signature:  #365
 Badge No.: 365
 Date: 04/18/2020