

SOLN* 303
4/24/20

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01845	Calib. Date:	04/24/2020
Certification File No.:	01846	Calib. No.:	00049
Linearity File No.:	01847	Cert. Date:	04/24/2020
Solution File No.:	01848	Cert. No.:	00036
Sequential File No.:	01848	Lin. Date:	04/24/2020
		Lin. No.:	00036
		Soln. Date:	04/24/2020
		Soln. No.:	00303
		File Date:	04/24/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0161

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:55D		
Control 1 EC	0.100%	10:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		
Control 2 EC	0.099%	10:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		
Control 3 EC	0.099%	10:58D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXK P2 - 444 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I Lent 7045

Badge No.: 7045

Date: 04/24/2020