

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01845 Calib. Date: 04/24/2020 Calib. No.: 00049  
Certification File No.: 01846 Cert. Date: 04/24/2020 Cert. No.: 00036  
Linearity File No.: 01847 Lin. Date: 04/24/2020 Lin. No.: 00036  
Solution File No.: 01852 Soln. Date: 05/21/2020 Soln. No.: 00304  
Sequential File No.: 01852 File Date: 05/21/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 02/19/2022  
Solution Control Lot: 20090 Bottle No.: 1309

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:10D		
Control 1 EC	0.099%	12:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.097%	12:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.098%	12:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: \_\_\_\_\_

Badge No.: 1293

Date: 05/21/2020