

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02859 Calib. Date: 04/03/2020 Calib. No.: 00037
Certification File No.: 02860 Cert. Date: 04/03/2020 Cert. No.: 00027
Linearity File No.: 02861 Lin. Date: 04/03/2020 Lin. No.: 00027
Solution File No.: 02887 Soln. Date: 06/09/2020 Soln. No.: 00357
Sequential File No.: 02887 File Date: 06/09/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/09/2021
Solution Control Lot: 19240 Bottle No.: 0105

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:47D		
Control 1 EC	0.099%	17:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:48D		
Control 2 EC	0.098%	17:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:50D		
Control 3 EC	0.098%	17:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 06/09/2020