

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02859	Calib. Date: 04/03/2020
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/03/2020
Solution File No.:	02893	Cert. No.: 00027
Sequential File No.:	02893	Lin. Date: 04/03/2020
		Lin. No.: 00027
		Soln. Date: 06/26/2020
		Soln. No.: 00358
		File Date: 06/26/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0514

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.099%	11:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.097%	11:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 3 EC	0.098%	11:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA	First Name: PASCAL	MI: M
Signature: <u></u>	# 365	Badge No.: 365
		Date: 06/26/2020