

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02859	Calib. Date: 04/03/2020
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/03/2020
Solution File No.:	02898	Cert. No.: 00027
Sequential File No.:	02898	Lin. Date: 04/03/2020
		Lin. No.: 00027
		Soln. Date: 07/04/2020
		Soln. No.: 00359
		File Date: 07/04/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 02/19/2022
Solution Control Lot:	20090		Bottle No.: 0452

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:40D		
Control 1 EC	0.097%	08:41D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		
Control 2 EC	0.097%	08:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42D		
Control 3 EC	0.098%	08:43D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:44D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 07/04/2020