

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01845	Calib. Date: 04/24/2020
Certification File No.:	01846	Calib. No.: 00049
Linearity File No.:	01847	Cert. Date: 04/24/2020
Solution File No.:	01863	Cert. No.: 00036
Sequential File No.:	01863	Lin. Date: 04/24/2020
		Lin. No.: 00036
		Soln. Date: 07/31/2020
		Soln. No.: 00307
		File Date: 07/31/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0051

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:26D		
Control 1 EC	0.099%	11:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 2 EC	0.098%	11:27D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.097%	11:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GAMBARDILLA

First Name: PASCAL

MI: M

Signature: _____



Badge No.: 365

Date: 07/31/2020