

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	04/03/2020
Calibration File No.:	02859	Calib. No.:	00037
Certification File No.:	02860	Cert. Date:	04/03/2020
Linearity File No.:	02861	Lin. Date:	04/03/2020
Solution File No.:	02923	Soln. Date:	08/14/2020
Sequential File No.:	02923	File Date:	08/14/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	02/19/2022
Solution Control Lot:	20090			Bottle No.:	0745

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:30D		
Control 1 EC	0.096%	23:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	23:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		
Control 2 EC	0.097%	23:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:32D		
Control 3 EC	0.097%	23:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 08/14/2020