

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02859	Calib. Date: 04/03/2020
Certification File No.:	02860	Calib. No.: 00037
Linearity File No.:	02861	Cert. Date: 04/03/2020
Solution File No.:	02926	Cert. No.: 00027
Sequential File No.:	02926	Lin. Date: 04/03/2020
		Lin. No.: 00027
		Soln. Date: 08/21/2020
		Soln. No.: 00364
		File Date: 08/21/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 02/19/2022
Solution Control Lot:	20090		Bottle No.: 0755

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:27D		
Control 1 EC	0.096%	11:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 2 EC	0.096%	11:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30D		
Control 3 EC	0.097%	11:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 08/21/2020