

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01870 Calib. Date: 08/25/2020 Calib. No.: 00050
Certification File No.: 01871 Cert. Date: 08/25/2020 Cert. No.: 00037
Linearity File No.: 01872 Lin. Date: 08/25/2020 Lin. No.: 00037
Solution File No.: 01873 Soln. Date: 08/25/2020 Soln. No.: 00309
Sequential File No.: 01873 File Date: 08/25/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 02/19/2022
Solution Control Lot: 20090 Bottle No.: 0108

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:21D		
Control 1 EC	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:23D		
Control 2 EC	0.100%	14:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24D		
Control 3 EC	0.098%	14:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-026 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I Dent 7045

Badge No.: 7045

Date: 08/25/2020