

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02929	Calib. Date: 08/25/2020
Certification File No.:	02930	Calib. No.: 00038
Linearity File No.:	02931	Cert. Date: 08/25/2020
Solution File No.:	02932	Cert. No.: 00028
Sequential File No.:	02932	Lin. Date: 08/25/2020
		Lin. No.: 00028
		Soln. Date: 08/25/2020
		Soln. No.: 00365
		File Date: 08/25/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0339

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:43D		
Control 1 EC	0.101%	10:43D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:43D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		
Control 2 EC	0.100%	10:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45D		
Control 3 EC	0.100%	10:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-025 (DL)

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature: Tpr J Lutz 7045 Badge No.: 7045 Date: 08/25/2020