

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01870	Calib. Date: 08/25/2020
Certification File No.:	01871	Calib. No.: 00050
Linearity File No.:	01872	Cert. Date: 08/25/2020
Solution File No.:	01876	Cert. No.: 00037
Sequential File No.:	01876	Lin. Date: 08/25/2020
		Lin. No.: 00037
		Soln. Date: 09/18/2020
		Soln. No.: 00310
		File Date: 09/18/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:38D		
Control 1 EC	0.099%	11:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40D		
Control 2 EC	0.098%	11:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41D		
Control 3 EC	0.098%	11:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Badge No.: 365

Date: 09/18/2020