

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02929	Calib. Date: 08/25/2020
Certification File No.:	02930	Calib. No.: 00038
Linearity File No.:	02931	Cert. Date: 08/25/2020
Solution File No.:	02951	Cert. No.: 00028
Sequential File No.:	02951	Lin. Date: 08/25/2020
		Lin. No.: 00028
		Soln. Date: 09/19/2020
		Soln. No.: 00367
		File Date: 09/19/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 1313

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:06D		
Control 1 EC	0.096%	10:06D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:06D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		
Control 2 EC	0.096%	10:07D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:07D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		
Control 3 EC	0.097%	10:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:



Badge No.: 365

Date: 09/19/2020