

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753		2. Contact: Robert Rankin 3. Phone Number: 732-349-0150	4. Date: 2/28/20 5. County: Ocean
6. Alcotest Instrument Serial Number: ARAJ-0075	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

Complaint received of the instrument experiencing multiple control test failures.

10. Comments:

Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed multiple control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Inc for evaluation.

Last Known Sequential File #: 2856


☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045
Name & Badge Number (Print)


Signature

2-28-20
Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753		2. Contact: Robert Rankin	4. Date: 4/3/20
		3. Phone Number: 732-349-0150	5. County: Ocean
6. Alcotest Instrument Serial Number: ARAJ-0075	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:

The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:

See Draeger Return and Repair form.

Last Known Sequential File #: NA

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.


☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045

Name & Badge Number (Print)


Signature

7045 4-3-20
Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Toms River Twp Police Department, Toms River, NJ

Date Received: 3/12/2020

Date given to service: 3/12/2020

Carrier: ☐ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR AJ-0075

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

FUEL CELL = ARMM-1266

CAL W/QC AND OPS CHECK

Service Technician

Date: 03/26/2020

Delivery receipt



Customer no.
150059761

Number of report
305000291

Please reference on inquiries

Date of receiving Delivery receipt date
/ / 03/27/2020

Consignee, 150059763

TOWNSHIP OF TOMS RIVER POLICE DEPAR
ATTN: SGT. ROBERT RANKIN
PO#ARAJ-0075
255 OAK AVE
TOMS RIVER NJ 08753-3383

Customer, 150059761

TOWNSHIP OF TOMS RIVER
ATTN: SGT. ROBERT RANKIN
PO#ARAJ-0075
255 OAK AVE
TOMS RIVER NJ 08753-0876

Your order

Date of order:
03/09/2020
Purch.ord.no.:
ARAJ-0075

Branch text

Draeger contact person

Service Support-IDT
800-437-2437 OPT 5
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030320499	ARAJ-0075
	REPAIR/ARAJ-0075			

Draeger Inc.
Our Tax ID: 23-1699096
3135 Quarry Road; Telford, PA 18969
An Equal Opportunity Employer M / F / V / H
Telephone 800-437-2437
<http://www.draeger.com>

Remit to:
LOCKBOX (Standard USPS)
Draeger, Inc.
PO Box 13369
Newark, New Jersey
07101-3362

Remit to:
LOCKBOX (Overnight)
FIS Lockbox Processing
Lockbox #13369
100 Grove Road
Suite E
West Deptford, NJ 08066

Remit US Wire Transfers to:
Account Name: Draeger Inc.
Account Number: 00-494-936
Transit Routing: 021001033
SWIFT: BKTRUS33
Deutsche Bank Trust Company Americas
60 Wall Street 25th Fl, New York, NY 10005

Service note

Customer no.
150059761

Number of report
305000291

Date of receiving

Please reference on inquiries

Customer
TOWNSHIP OF TOMS RIVER
255 OAK AVE
TOMS RIVER NJ 08753-0876

Consignee
TOWNSHIP OF TOMS RIVER POLICE DEPAR
ATTN: CHIEF OF POLICE
255 OAK AVE
TOMS RIVER NJ 08753-3383

Your order

Date of order:

03/09/2020

Your reference:

ARAJ-0075

NL-Text

Your contact person

Service Support-IDT

800-437-2437 OPT 5

SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 142696469
		REPAIR/ARAJ-0075
		Eq.No. : 1030320499 SN: ARAJ-0075
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		REPAIR/ARAJ-0075
1.0	H	R001 Repair time - normal hours
1	EA	6808374 Geared Engine, A7110/A9510
1	EA	6808455 Sensor Alcotest B