

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753</b>		2. Contact: <b>Robert Rankin</b> 3. Phone Number: <b>732-349-0150</b>	4. Date: <b>2/28/20</b> 5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARWF-0394</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:

**Complaint received of the instrument experiencing multiple control test failures.**

10. Comments:

**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed multiple control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Inc for evaluation.**

Last Known Sequential File #: 1838

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☒ 12. The above Instrument/Component is placed out of service pending further evaluation.

☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045  
Name & Badge Number (Print)

*Tpr I Lutz* 7045 2-28-20  
Signature Date

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753</b>		2. Contact: <b>Robert Rankin</b> 3. Phone Number: <b>732-349-0150</b>	4. Date: <b>4/24/20</b> 5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARWF-0394</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:

**The alcotest instrument was returned from outside evaluation and placed back in service.**

10. Comments:

**See Draeger Return and Repair form.**

**Last Known Sequential File #: NA**

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<u>Tpr. I Dennis J. Lutz #7045</u> Name & Badge Number (Print)	<u>Tpr I Lutz 7045</u> Signature	<u>4-24-20</u> Date
---	-------------------------------------	------------------------

# RETURN AND REPAIR FORM

**Dräger**

## Customer Information

B:

S:

Company Name: Toms River Twp Police Department, Toms River, NJ

Date Received: 03/12/2020

Date given to service: 03/12/2020

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR WF-0394

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

FUEL CELL = ARMH-0603

CAL W/QC AND OPS CHECK

Service Technician: 

Date: 04/09/2020

## Delivery receipt



Customer no.  
150059761

Number of report  
305000285  
Please reference on inquiries

Date of receiving    Delivery receipt date  
/    /    04/10/2020

Consignee, 150059763  
**TOWNSHIP OF TOMS RIVER POLICE DEPAR**  
**ATTN: SGT. ROBERT RANKIN**  
**PO#ARWF-0394**  
**255 OAK AVE**  
**TOMS RIVER NJ    08753-3383**

Customer, 150059761  
TOWNSHIP OF TOMS RIVER  
ATTN: SGT. ROBERT RANKIN  
PO#ARWF-0394  
255 OAK AVE  
TOMS RIVER NJ    08753-0876

### Your order

Date of order:  
03/09/2020  
Purch.ord.no.:  
ARWF-0394

### Branch text

### Draeger contact person

Service Support-IDT  
800-437-2437 OPT 5  
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
	ALCOTEST 7110 MK III USA	8314246	1030320503	ARWF-0394
	REPAIR/ARWF-0394			

**Draeger Inc.**  
Our Tax ID: 23-1699096  
3135 Quarry Road; Telford, PA 18969  
An Equal Opportunity Employer M / F / V / H  
Telephone 800-437-2437  
<http://www.draeger.com>

Remit to:  
LOCKBOX (Standard USPS)  
Draeger, Inc.  
PO Box 13369  
Newark, New Jersey  
07101-3362

Remit to:  
LOCKBOX (Overnight)  
FIS Lockbox Processing  
Lockbox #13369  
100 Grove Road  
Suite E  
West Deptford, NJ 08066

Remit US Wire Transfers to:  
Account Name: Draeger Inc.  
Account Number: 00-494-936  
Transit Routing: 021001033  
SWIFT: BKTRUS33  
Deutsche Bank Trust Company Americas  
60 Wall Street 25th Fl, New York, NY 10005

# Service note

Customer no. 150059761  
Number of report 305000285  
Date of receiving  
Please reference on inquiries

Customer  
TOWNSHIP OF TOMS RIVER  
255 OAK AVE  
TOMS RIVER NJ 08753-0876

Consignee  
TOWNSHIP OF TOMS RIVER POLICE DEPAR  
ATTN: CHIEF OF POLICE  
255 OAK AVE  
TOMS RIVER NJ 08753-3383

Your order  
Date of order:  
03/09/2020  
Your reference:  
ARWF-0394

NL-Text

Your contact person  
Service Support-IDT  
800-437-2437 OPT 5  
SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		<b>Service Order: 142696468</b>
		REPAIR/ARWF-0394
		Eq.No. : 1030320503 SN: ARWF-0394
		Material: 8314246 ALCOTEST 7110 MK III USA
		Inv.No. :
		Location:
		<b>REPAIR/ARWF-0394</b>
1.0	H	R001 Repair time - normal hours
1	EA	6808374 Geared Engine, A7110/A9510
1	EA	6808455 Sensor Alcotest B