

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02929	Calib. Date: 08/25/2020
Certification File No.:	02930	Calib. No.: 00038
Linearity File No.:	02931	Cert. Date: 08/25/2020
Solution File No.:	02963	Cert. No.: 00028
Sequential File No.:	02963	Lin. Date: 08/25/2020
		Lin. No.: 00028
		Soln. Date: 10/09/2020
		Soln. No.: 00369
		File Date: 10/09/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0340

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:37D		
Control 1 EC	0.097%	12:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38D		
Control 2 EC	0.096%	12:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39D		
Control 3 EC	0.097%	12:40D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:40D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 10/09/2020