

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02929 Calib. Date: 08/25/2020 Calib. No.: 00038
Certification File No.: 02930 Cert. Date: 08/25/2020 Cert. No.: 00028
Linearity File No.: 02931 Lin. Date: 08/25/2020 Lin. No.: 00028
Solution File No.: 02970 Soln. Date: 10/22/2020 Soln. No.: 00370
Sequential File No.: 02970 File Date: 10/22/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0151

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:52D		
Control 1 EC	0.097%	12:53D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:53D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 2 EC	0.096%	12:54D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:54D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		
Control 3 EC	0.097%	12:56D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:56D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 10/22/2020