

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02929	Calib. Date: 08/25/2020
Certification File No.:	02930	Calib. No.: 00038
Linearity File No.:	02931	Cert. Date: 08/25/2020
Solution File No.:	02974	Cert. No.: 00028
Sequential File No.:	02974	Lin. Date: 08/25/2020
		Lin. No.: 00028
		Soln. Date: 10/26/2020
		Soln. No.: 00371
		File Date: 10/26/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0386
Control Solution %:	0.100%			Expires:	09/02/2022
Solution Control Lot:	20390			Bottle No.:	0210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.096%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.096%	07:38S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:38S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		
Control 3 EC	0.097%	07:39S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:39S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BARTOSHEK

First Name: PTL

MI: D

Signature: \_\_\_\_\_

*David J. Bartoshek* #316 TLPD

Badge No.: 316

Date: 10/26/2020