

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02929	Calib. Date: 08/25/2020
Certification File No.:	02930	Calib. No.: 00038
Linearity File No.:	02931	Cert. Date: 08/25/2020
Solution File No.:	02997	Cert. No.: 00028
Sequential File No.:	02997	Lin. Date: 08/25/2020
		Lin. No.: 00028
		Soln. Date: 12/24/2020
		Soln. No.: 00376
		File Date: 12/24/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWK S3-0386
Control Solution %:	0.100%		Expires: 09/02/2022
Solution Control Lot:	20390		Bottle No.: 0045

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:19S		
Control 1 EC	0.101%	11:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21S		
Control 2 EC	0.099%	11:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22S		
Control 3 EC	0.099%	11:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Badge No.: 365

Date: 12/24/2020