

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02929 Calib. Date: 08/25/2020 Calib. No.: 00038  
Certification File No.: 02930 Cert. Date: 08/25/2020 Cert. No.: 00028  
Linearity File No.: 02931 Lin. Date: 08/25/2020 Lin. No.: 00028  
Solution File No.: 03001 Soln. Date: 01/08/2021 Soln. No.: 00377  
Sequential File No.: 03001 File Date: 01/08/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 0213

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:36S		
Control 1 EC	0.101%	11:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38S		
Control 2 EC	0.099%	11:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39S		
Control 3 EC	0.099%	11:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature:  #399

Badge No.: 1399

Date: 01/08/2021