

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02929 Calib. Date: 08/25/2020 Calib. No.: 00038  
Certification File No.: 02930 Cert. Date: 08/25/2020 Cert. No.: 00028  
Linearity File No.: 02931 Lin. Date: 08/25/2020 Lin. No.: 00028  
Solution File No.: 03021 Soln. Date: 02/04/2021 Soln. No.: 00379  
Sequential File No.: 03021 File Date: 02/04/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 0099

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:07S		
Control 1 EC	0.101%	12:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		
Control 2 EC	0.099%	12:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09S		
Control 3 EC	0.099%	12:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: 

Badge No.: 336

Date: 02/04/2021



**ORIGINAL**