

ITEM 15
SOLN 380
2-12-21

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029
Solution File No.: 03030 Soln. Date: 02/12/2021 Soln. No.: 00380
Sequential File No.: 03030 File Date: 02/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0043

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:44S		
Control 1 EC	0.100%	10:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		
Control 2 EC	0.097%	10:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47S		
Control 3 EC	0.097%	10:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436 (DL)

Changed By:
Last Name: LUTZ First Name: DENNIS MI: J
Signature: Tpr I Lutz 7045 Badge No.: 7045
Date: 02/12/2021