

SOLN 381
2/27/21

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029
Solution File No.: 03034 Soln. Date: 02/27/2021 Soln. No.: 00381
Sequential File No.: 03034 File Date: 02/27/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0044

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:28S		
Control 1 EC	0.098%	11:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30S		
Control 2 EC	0.096%	11:30S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 3 EC	0.097%	11:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 02/27/2021