

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039  
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029  
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029  
Solution File No.: 03042 Soln. Date: 03/06/2021 Soln. No.: 00382  
Sequential File No.: 03042 File Date: 03/06/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 0042

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:07S		
Control 1 EC	0.100%	04:08S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:09S		
Control 2 EC	0.097%	04:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:10S		
Control 3 EC	0.098%	04:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: WESTHALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_

Badge No.: 306

Date: 03/06/2021