

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03027	Calib. Date: 02/12/2021
Certification File No.:	03028	Calib. No.: 00039
Linearity File No.:	03029	Cert. Date: 02/12/2021
Solution File No.:	03054	Cert. No.: 00029
Sequential File No.:	03054	Lin. Date: 02/12/2021
		Lin. No.: 00029
		Soln. Date: 03/16/2021
		Soln. No.: 00384
		File Date: 03/16/2021

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 09/02/2022
Solution Control Lot:	20390		Bottle No.: 0214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:55S		
Control 1 EC	0.097%	15:55S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	15:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:56S		
Control 2 EC	0.096%	15:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	15:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57S		
Control 3 EC	0.096%	15:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	15:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 03/16/2021