

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest-7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01891	Calib. Date:	02/12/2021
Certification File No.:	01892	Cert. No.:	00051
Linearity File No.:	01893	Cert. Date:	02/12/2021
Solution File No.:	01901	Lin. Date:	02/12/2021
Sequential File No.:	01901	Lin. No.:	00038
		Soln. Date:	03/18/2021
		Soln. No.:	00319
		File Date:	03/18/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	0041

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:26S		
Control 1 EC	0.099%	13:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27S		
Control 2 EC	0.096%	13:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		
Control 3 EC	0.096%	13:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 03/18/2021